



## Decision Matrix

### SWIVEL BATHER

#### General:

- Weight of client? Not exceed 250 lbs for the SB200 model and not to exceed 300 lbs for the SB400 model.
- Seat Width of client?
- Note: this device has been designed for ambulatory users ONLY. If the client cannot walk to the tub, turn around with their back to the device, reach back and grasp the armrests and sit down (rather than falling backwards into the device), WITHOUT ASSISTANCE, from a care giver then they are not a good candidate for the Swivel-Bather.
- Does the client have the upper body strength to lower self into swivel seat without dropping into the seat?
- To calculate the seat height of the Swivel-Bather, determine the outside tub wall height and add 6".